

## UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER OF PATENTS AND TRADEMARKS Washington, D.C. 20231 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	KET NO. CONFIRMATION NO.		
09/682,001	07/06/2001	Alexander Ganin	15-XZ-5881	8764		
23446 7	7590 12/11/2001					
MCANDREWS HELD & MALLOY, LTD			EXAMINER			
<b>SUITE 3400</b>	• • • • • • • • • • • • • • • • • • • •			CHURCH, CRAIG E		
CHICAGO, IL	. 60661		ART UNIT	PAPER NUMBER		
			2882	2.3		
			DATE MAILED: 12/11/2001	·		

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No. Applicant(s)					
Office Action Summary	Examiner		Group Art Unit			
-The MAILING DATE of this communication appears	on the cover sheet be	eneath the co	rrespondence ac	idress		
Peri d for Reply						
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO OF THIS COMMUNICATION.	EXPIRE 3	MONTH(S)	FROM THE MAIL	ING DATE		
<ul> <li>Extensions of time may be available under the provisions of 37 CFR 1.13 from the mailing date of this communication.</li> <li>If the period for reply specified above is less than thirty (30) days, a reply</li> <li>If NO period for reply is specified above, such period shall, by default, ex</li> <li>Failure to reply within the set or extended period for reply will, by statute,</li> </ul>	within the statutory minima pire SIX (6) MONTHS from	um of thirty (30) on the mailing date	days will be considered	ed timely.		
Status	,					
☐ Responsive to communication(s) filed on				<del></del> •		
☐ This action is FINAL.						
☐ Since this application is in condition for allowance except fo accordance with the practice under Ex parte Quayle, 1935 €			the merits is clos	sed in		
Disposition of Claims						
▼ Claim(s) 1-2	is/are pending in the application.					
Of the above claim(s)	is/are v	is/are withdrawn from consideration.				
☐ Claim(s)	is/are a	is/are allowed.				
☑ Claim(s) 1-2/	is/are re	is/are rejected.				
☐ Claim(s)	is/are objected to.					
☐ Claim(s)		are sub		or election		
Application Papers						
☐ See the attached Notice of Draftsperson's Patent Drawing F						
☐ The proposed drawing correction, filed on is ☐ approved ☐ disapproved.						
☐ The drawing(s) filed on is/are objected ☐ The specification is objected to by the Examiner.	to by the Examiner.					
☐ The oath or declaration is objected to by the Examiner.						
Priority under 35 U.S.C. § 119 (a)-(d)						
☐ Acknowledgment is made of a claim for foreign priority unde	or 35 U.S.C. & 11 9/a\-/	'd)				
<ul> <li>□ All □ Some* □ None of the CERTIFIED copies of the</li> <li>□ received.</li> </ul>	priority documents ha	ve been				
<ul> <li>received in Application No. (Series Code/Serial Number)</li> <li>received in this national stage application from the International</li> </ul>			<del></del> •			
*Certified copies not received:			·			
Attachment(s)						
☑ Information Disclosure Statement(s), PTO-1449, Paper No(	s) 🗆 Ir	terview Sumn	nary, PTO-413			
☑ Notice of Reference(s) Cited, PTO-892	•		nal Patent Applicat	ion, PTO-152		
☐ Notice of Draftsperson's Patent Drawing Revi w, PTO-948						
Office Action Summary						

est with

Serial No. 682,001 Art Unit 2882

The following is a quotation of the first paragraph of 35 U.S.C. § 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

The specification is objected to under 35 U.S.C. § 112, first paragraph, as failing to provide an adequate written description. The meaning and significance of "servo-tomo function" are not defined.

Claims 1-9 are rejected under 35 U.S.C. § 112, first paragraph, for the reasons set forth in the objection to the specification.

Claims 1-9 are rejected under 35 U.S.C. § 112, first paragraph, for the reasons set forth in the objection to the specification. The meaning and significance of "servo-tomo function" are not defined.

The following is a quotation of 35 U.S.C. § 103 which forms the basis for all obviousness rejections set forth in this Office action:

A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention

Serial No. Art Unit

682,001 2882

was made.

Subject matter developed by another person, which qualifies as prior art only under subsection (f) or (g) of section 102 of this title, shall not preclude patentability under this section where the subject matter and the claimed invention were, at the time the invention was made, owned by the same person or subject to an obligation of assignment to the same person.

Claims 1-9 are rejected under 35 U.S.C. § 103 as being unpatentable over Khutoryansky et al (5734694) cited by applicant. Lines 50 of column 4 to 38 of column 6 explain operation of the disclosed tomographic system which includes presetting parameters such as angle, fulcrum (focal plane) and focal plane thickness. Switches >/< specify clockwise and counterclockwise scanning, and it would have been obvious to one of ordinary skill in the art at the time the invention was made to operate them in sequence since the source scanning in either direction is limited.

Claims 6, 7 and 10-21 are rejected under 35 U.S.C. § 103 as being unpatentable over Khutoryansky et al (5734694) in view of either Lin et al (6222902) or Kruger (4577222). Lin and Kruger teach real time digital display of multiple tomographic slices, and it would have been obvious to one of ordinary skill in the art at the time the invention was made to equip the Khutoryansky with such means to speed up medical diagnosis.

Any inquiry concerning this communication should be directed to Examiner Church at telephone number (703) 308-861

CRAIG E. CHURCH Senior Examiner ART UNIT 2882